

Certificate of Calibration

Issued By Trescal Ltd



0013

Date of Issue: 17 January 2018

Certificate Number: 2622520004

Page 1 of 3

Trescal Ltd
Park Gate Close, Bredbury Park Way
Bredbury, Stockport,
SK6 2SL, UK
Tel: +44(0) 161 406 7878
Fax: +44(0) 161 406 7979
Email: calibration.manchester@trescal.com

APPROVED SIGNATORY

Pedro Mendes

Customer:

Pyser -SGI Limited, Fircroft Way
Edenbridge, , Kent, TN8 6HA

Equipment Details

Description: Stage Micrometer

Manufacturer: Pyser-SGI

Type No: PS8

Date of Receipt: 11/Jan/2018

Range: 1.0 x 0.01 mm

Order No: 098684

Serial No: CS5225

Our Reference: 00459206

Calibrated By: Sam Hallas

Date: 17/Jan/2018

Calibration Summary

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Ambient Conditions Temperature: 20°C ± 2°C

Date of next calibration: 17/Jan/2020

The results given within this certificate only relate to the item calibrated. The uncertainty limits quoted refer to the measured values only, with no account being taken of the instruments ability to maintain its calibration. The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$ providing a confidence level of approximately 95%. The uncertainty evaluation has been derived from EA-4/02 M:2013 "Evaluation of the Uncertainty of Measurement In Calibration".

EMS 00004-27-Nov2017